


# PRELIMINARY RESULTS FROM THE MICROSCALE COMBUSTION CALORIMETER ROUND ROBIN



Richard Walters FAA

Natallia Safronava TAMI

# ASTM E691-09

## Conducting an Inter-Laboratory Study to Determine the Precision of a Method

- **Repeatability**
  - An action, event, or other thing that is done again
- **Reproducibility**
  - Create something very similar to (something else) in a different medium or context

# Microscale Combustion Calorimeter

**Developed by FAA Fire Safety Team**

**Small scale flammability test**

**Quantitative results**


**HRR vs Temperature**

**10-15 minute test**

**Sample size of 5mg**



# MCC Schematic



# Measured Values



**Sample Weight (mg)**


**Char Yield (%)**

**Peak Heat Release Rate (W/g)**

**Total Heat Released (kJ/g)**

**Peak Heat Release Temperature (°C)**

**Heat Release Capacity (J/g-K)**



# Heat Release Capacity Measurements & Theory

## Heat Release Rate

$$\dot{Q}_c(t) = \frac{E\Delta O_2}{m_o} = h_{c,v}^o(t) \left[ \frac{-1}{m_o} \frac{dm(t)}{dt} \right]$$

## Peak Heat Release Rate

$$\dot{Q}_c^{\max} = \left[ \frac{E\Delta O_2}{m_o} \right]_{\max} = h_{c,v}^o \left[ \frac{-1}{m_o} \frac{dm}{dt} \right]_{\max} = h_{c,v}^o \frac{\beta(1-\mu)E_a}{eR T_p^2}$$

## Heat Release Capacity


$$\eta_c = \frac{\dot{Q}_c^{\max}(\beta)}{\beta} = \frac{h_c^o(1-\mu)E_a}{eR T_p^2}$$

# Sample Preparation

- 5 samples selected
- Hole punch used to make consistent sample sizes
- Hole size depended on film thickness and density
- Samples ranged from 3.2 mg to 6.2 mg




# Samples “as Received”



<b>Acrylic</b>	<b>3.18 mg</b>
<b>Polystyrene</b>	<b>4.55 mg</b>
<b>Polypropylene</b>	<b>3.44 mg</b>
<b>Polycarbonate</b>	<b>3.47 mg</b>
<b>Polyphenylsulfone</b>	<b>6.20 mg</b>




# Sample Heat Release Rates




# Interlaboratory Study

- **3 manufacturers of the MCC**
  - **Manufacturer A**
  - **Manufacturer B**
  - **Manufacturer C**
- **Samples sent to 18 labs**
  - **Several labs could not participate due to equipment not working**
- **Data received from 11 labs so far**
  - **1 lab rejected due to bad O<sub>2</sub> sensor**
  - **All raw data was re-analyzed and interpreted by me**


# Acrylic




# Polystyrene




# Polypropylene




# Polycarbonate



# Polyphenylsulfone




# Sample Weight







# Peak Heat Release Rate




# Heat Release Capacity




# Total Heat Release



# Temperature at Peak HRR




# Char Yield




# Lab Comparison – Peak HRR

## AVG




## REF




# Lab Comparison – HR Capacity

## AVG




## REF




# Lab Comparison – Total HR

## AVG





## REF







# Lab Comparison – Temperature



- Reference temperature obtained from calibrated thermogravimetric analyzer (TGA) measurements

# Lab Comparison – Char Yield



- Char values show a lot of scatter due to different maximum temperature settings

# Sources of Error

- **Heat Release Rate**
  - **Bad O<sub>2</sub> sensor calibration**
  - **Leak in system**
  - **Sample size**
  - **Bad weight measurements**
  - **Flow meter/controllers out of calibration**
- **Temperature**
  - **Thermal lag**
  - **DAQ board calibration**
  - **Time shift**

# Temperature Calibration

- **Why is it needed?**
  - **Reproducibility between laboratories**
- **What does it do?**
  - **Corrects for thermal lag when making dynamic measurements**

# Temperature Calibration - Metals



**Indium: 156.6°C**

**Tin: 231.9°C**

**Zinc: 419.5°C**

**Antimony: 630.7°C**

**Aluminum: 660.4°C**



**Sodium Chloride: 801°C**

**Silver: 961°C**

**Calibration Kit**  
**99.999% Pure**  
**New Dedicated Pans**




# Temperature Calibration – Metal Foils




- **Linear deviation**
- **Largest deviation at high temperature**
- **“Melting” observed with little to no discoloration**
- **Correlates with TGA**

# MCC vs TGA



**MCC T calibration brings measured temperature within 4 degrees of TGA and 1 degree of other MCC**

# Repeatability & Reproducibility



- **13 Tests**
- **3 Laboratories**
- **4 Operators**



# Conclusions & Future Work

- **More labs will be added to study as the data is received**
- **Temperature calibration will be implemented**
- **Comprehensive users guide will be available soon**
- **More explicit instructions will be given to the participants in the future**
- **One of the manufacturers has issues that need to be dealt with**
- **Inter-laboratory study – round 2**

# **WANTED: More Participants**

## **Contact Information:**

**Richard Walters**

**FAA William J Hughes Technical Center**

**Fire Safety Team AJP-6320**

**Atlantic City International Airport, NJ 08405**

**609 485-4328**

**[richard.walters@faa.gov](mailto:richard.walters@faa.gov)**