



875 WATTS  
120 VOLTS  
60 WATTS/SQ.IN.

(B) 117 W/in<sup>2</sup> WATT DENSITY OF INTERNAL RESISTANCE WIRE

REVISIONS			
SYM	DESCRIPTION	DATE	APPROVED
A	1 1/8" ± 1/4, WAS 34° ± 1/4" ADDED COLD END LOCATION	11-8-07	P. ZIEG
B	ADDED WATT DENSITY, INCONEL WAS INCOLOY, 60 WATTS/SQ.IN. WAS 50 WATTS/SQ.IN.	1/15/08	P. ZIEG

UNLESS OTHERWISE SPECIFIED  
DIMENSIONS ARE IN INCHES  
TOLERANCES:  
FRACTIONS ±  
DECIMALS ±  
ANGLES ±  
REMOVE ALL BURRS AND SHARP EDGES

**NEWPORT SCIENTIFIC, INC.**  
8246-E SANDY COURT  
JESSUP, MD 20794  
NEWPORT-SCIENTIFIC.COM  
sales@newport-scientific.com

DO NOT SCALE THIS DRAWING

BY NAME DATE FINISH:

CHECKED P. ZIEG 1/15/08

ENGR/DSGN

APPROVED

MATERIAL

FINISH:

DO NOT SCALE THIS DRAWING

DO NOT SCALE THIS DRAWING

DO NOT SCALE THIS DRAWING

DO NOT SCALE THIS DRAWING

DO NOT SCALE THIS DRAWING



**FORMED TUBULAR HEATER**

CODE IDENT. NO. 59505

SIZE

PART NO. 68086038700  
DRAWING NO. 38768086